



Tera-PEEK

PEEK BASED POSTERIOR LUMBAR SYSTEM



The Tera-PEEK system footprint **is intended for a posterior approach with a large graft window volume. Tera-PEEK is available in a larger variety of lengths and widths than competitive systems in up to 16° of lordosis.**

VALUE FOR SURGEONS AND PATIENTS

The Tera-PEEK system is provided in multiple footprints and is available in PEEK with Tantalum Markers.

Endplates provide a larger surface of implant/bone contact compared to competitive systems.


Implants are only provided sterile packaged, reducing the risks of delays or canceled surgeries due to lack of availability.

INDICATIONS FOR USE

The Tera-PEEK system is **indicated for intervertebral body fusion of the spine in skeletally mature patients who have had at least six months of non-operative treatment.** The system is designed for use with allogenic bone graft comprised of cancellous and/or corticocancellous bone graft and/or autograft to facilitate fusion.

The Tera-PEEK system is **intended for use at either one level or two contiguous levels in the lumbar spine, from L2 to S1, for the treatment of degenerative disc disease (DDD) with up to Grade I spondylolisthesis.** DDD is defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies.

The Tera-PEEK system is **intended to be used with a supplemental spinal fixation system** cleared for use in the lumbosacral spine.



Tera-PEEK (PLIF) PEEK BASED POSTERIOR LUMBAR SYSTEM

SIZES

LENGTH	20mm – 35mm
WIDTH	8mm – 14mm
HEIGHT	6mm – 17mm
LORDOSIS	0°, 8°, 16°