



ALIGN-N

STAND ALONE ANTERIOR LUMBAR
SYSTEM WITH NAIL FIXATION

*Align-N includes two nail anchor options and two screw options in various footprints to **provide maximum flexibility in meeting all patient needs.** Align-N can be used as a stand-alone device when two titanium alloy fixed angle screws are used to promote long term stability.*

*Note: The system must be used with supplemental fixation when using nail anchors. The Align-N system's design and porous structure **improves adhesion to vertebral body endplates to support bone in-growth** when used with allogenic bone graft and/or autograft to facilitate fusion. As a stand-alone device, the Align-N system is **implanted through a single incision - promoting less trauma for the patient** and saving valuable time for the surgical team.*



ANTERIOR COLUMN ZERO PROFILE

Reduces or eliminates potential for great vessel interference

ENHANCED POROUS STRUCTURE

Increases bone in-growth to endplate

NAIL ANCHORS

Simple insertion with pre-loaded nail guide

LOCKING PLATE

Provides visual & audible confirmation of fully seated nail anchors and screws to prevent backout with break-off feature

ENHANCED POROUS STRUCTURE

Increases bone in-growth to endplate



S U R G I C A L

844.228.4890 • support@AcuitySurgical.com • AcuitySurgical.com

ALIGN-N

STAND ALONE ANTERIOR LUMBAR SYSTEM WITH NAIL FIXATION

INDICATIONS FOR USE

Align-N is indicated for intervertebral body fusion of the spine in skeletally mature patients who have had at least six months of non-operative treatment. The device systems are designed for use with allogenic bone graft comprised of cancellous and/or corticocancellous bone graft and/or autograft to facilitate fusion. One device is used per intervertebral body space.

Align-N is intended for use at either one level or two contiguous levels in the lumbar spine, from L2 to S1, for the treatment of degenerative disc disease (DDD) with up to Grade I spondylolisthesis. DDD is defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies.

Align-N 2-screw anterior cages may be used as a stand alone device only when two (2) vertebral body bone screws are used. If the physician chooses to use Align-N with fewer than two (2) screws in the two medial fixation holes with one inferior and one superior screw trajectory, or utilizes the nail anchors, then an additional supplemental spinal fixation system cleared for use in the lumbosacral spine must be used.



Nail Anchor
Diameter: 5.0 mm
Length: 20 | 25 | 30 mm



Cover Plate



Screws
Diameter: 5.0 | 5.5 mm
Length: 20 | 25 | 30 mm



Footprints

XS	32 x 21
S	32 x 24
L	36 x 26
XL	38 x 30

All footprints available in 4-screw configurations

UNITARY TITANIUM STANDARD SIZES	
Lordosis	Height
25°	- - - - - 15 16 17 18 19 20 21 22
20°	- - - - - 15 16 17 18 19 20 21 22
16°	- 11 12 13 14 15 16 17 18 19 20 21 22
12°	10 11 12 13 14 15 16 17 18 19 20 21 22
7°	10 11 12 13 14 15 16 17 18 19 20 21 22



Align-N Surgical Technique Guide is available upon request

844.228.4890 • support@AcuitySurgical.com • AcuitySurgical.com

LBL-AS202605 Rev A.01